

Veratox[®] for Aflatoxin M₁ Procedure

Note: Please read kit instructions completely before performing test.

1. Add 250 µL of controls and samples to their respective red-marked mixing wells.
2. Transfer 100 µL from the mixing wells to the antibody coated wells.
3. Place microwell holder on an automatic plate shaker and allow continuous shaking for 20 minutes at room temperature.
4. Dump liquid from antibody wells. Wash wells thoroughly with diluted wash buffer, repeat wash step 5 times. Tap out liquid on absorbent paper.
5. Transfer 100 µL of conjugate from the reagent boat to the wells.
6. Place microwell holder on an automatic plate shaker and allow continuous shaking for 10 minutes at room temperature.
7. Dump liquid from antibody wells. Wash wells thoroughly with diluted wash buffer, repeat wash step 5 times. Tap out liquid on absorbent paper.
8. Transfer 100 µL of substrate from the reagent boat to the antibody wells.
9. Place microwell holder on an automatic plate shaker and allow continuous shaking for 15 minutes at room temperature.
10. Transfer 100 µL of red stop solution from the reagent boat to the wells. Mix by sliding microwell holder back and forth on a flat surface.
11. Read results using a microwell reader with a 650 nm filter.



Product Specifications

Format: Competitive ELISA

Controls provided: 0, 5, 15, 30, 60 and 100 ppt aflatoxin M₁

Limit of detection: 4.3 ppt

Range of quantification: 5–100 ppt

Test time: 45 minutes

Ordering Information

8019 Veratox for Aflatoxin M₁

To place an order, or for more information, contact your local Neogen Europe representative, or visit our web site at www.neogeneurope.com

Neogen Europe Ltd. • The Dairy School • Auchincruive, Ayr, Scotland • KA6 5HU
Tel: +44 (0)1292 525610 • Fax: +44 (0)1292 525602
www.neogeneurope.com • info_uk@neogeneurope.com

© Neogen Corporation, 2013. Neogen and Veratox are registered trademarks of Neogen Corporation, Lansing, Michigan.

NE7219-1213